

4700 FESEM Standard Conditions

These are the conditions that the S4700 should be left in when you finish your session.

Stage	Home Position
Condenser Lens 1	5
Detector	MIXED
Operating mode	Normal
Specimen Bias Voltage	off
kV	15
Degas switch	Heat
Computer	On, user logged off
Ip knob	Ip3